



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Pelican Wire Company Inc.**  
**3650 Shaw Blvd.**  
**Naples, FL 34117**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 13 October 2026

Certificate Number: AC-3138



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Pelican Wire Company Inc.**  
3650 Shaw Blvd.  
Naples, FL 34117  
William L. Willett Jr. (239) 597-8555

**CALIBRATION**

Valid to: **October 13, 2026**

Certificate Number: **AC-3138**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance - Measure	(1 to 9) $\Omega$	0.018 $\Omega$	Keysight 34420A Multimeter
	(10 to 90) $\Omega$	0.29 $\Omega$	
	(100 to 900) $\Omega$	2.5 $\Omega$	
	(1 to 9) k $\Omega$	25 $\Omega$	
	(10 to 100) k $\Omega$	0.27 k $\Omega$	

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thermocouples	(93 to 148) $^{\circ}\text{C}$		Comparison to Fluke 5624 PRT (System 1)
	Type E	0.32 $^{\circ}\text{C}$	
	Type J	0.36 $^{\circ}\text{C}$	
	Type K	0.44 $^{\circ}\text{C}$	
	Type N	0.37 $^{\circ}\text{C}$	
Thermocouples	Type T	0.32 $^{\circ}\text{C}$	Comparison to Fluke 5624 PRT (System 1)
	(149 to 426) $^{\circ}\text{C}$		
	Type E	0.44 $^{\circ}\text{C}$	
	Type J	0.47 $^{\circ}\text{C}$	
	Type K	0.53 $^{\circ}\text{C}$	
	Type N	0.48 $^{\circ}\text{C}$	
	Type T	0.44 $^{\circ}\text{C}$	
	(426 to 649) $^{\circ}\text{C}$		
	Type E	0.59 $^{\circ}\text{C}$	
	Type J	0.61 $^{\circ}\text{C}$	
Type K	0.66 $^{\circ}\text{C}$		
Type N	0.62 $^{\circ}\text{C}$		

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thermocouples	(538 to 1 200) °C Type E Type J Type K Type N	0.78 °C 0.79 °C 0.83 °C 0.8 °C	Comparison to Fluke 5650 Type S Thermocouple Standard (System 1 HT)
Thermocouples	(93 to 148) °C Type E Type J Type K Type N Type T (149 to 426) °C Type E Type J Type K Type N Type T (426 to 649) °C Type E Type J Type K Type N	0.4 °C 0.44 °C 0.51 °C 0.44 °C 0.4 °C 0.5 °C 0.53 °C 0.59 °C 0.54 °C 0.5 °C 0.64 °C 0.66 °C 0.71 °C 0.66 °C	Comparison to Fluke 5650 Type S Thermocouple Standard (System 2)
Thermocouples	(-40 to 93) °C Type E Type J Type K Type N Type T	0.27 °C 0.31 °C 0.4 °C 0.32 °C 0.27 °C	Comparison to Fluke 5628 PRT (System 3)

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-3138.



Jason Stine, Vice President